

CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A12345D
NOMINAL SIZE IN (ul)	200
MANUFACTURER	Brand - 8 Ch.
DEPARTMENT	Your Dept.
LOCATION	Your Company
CONTACT	Your Contact
RESTRICTIONS	N/A
CALIBRATION DATE	30-Sep-2008
CALIBRATED BY	DAB
CALIBRATION DUE	Mar-2009
DATA TYPE	AS FOUND
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.0
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE mmHg	30.0
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HYGROMETER SERIAL #	1234
CALIBRATION DUE	Jan-2009
BALANCE ID	Brand
BALANCE MODEL	ABC123
BALANCE SERIAL #	ABCD1234
CALIBRATION DUE	Jan-2009
WEIGHT SET SERIAL #	ABC123
CALIBRATION DUE	Aug-2009
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	NO
PREVENTIVE MAINT.	NO
INSTRUMENT STATUS	PASS

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME ul	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE:AS FOUND
50	3.00	0.00150	3.00		
150	3.00	0.00450	3.00		
200	3.00	0.00600	3.00		

AS FOUND CALIBRATION DATA CHANNEL 1							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05045	0.15035	0.19959	STATISTICAL ANALYSIS
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010	+/- ml (Guardband)	0.00129	0.00429	0.00500	
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959	INACCURACY %	0.90008	0.23142	0.20320	
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909	IMPRECISION %	0.19881	0.06671	0.25126	
							Proc. Uncert.(k=2)	0.00021	0.00021	0.00100	
CALIBRATION RESULTS											
ACCURACY							PASS	PASS	PASS		
PRECISION							PASS	PASS	PASS		
Within k=2 95% Confidence							YES	YES	YES		
AS FOUND CALIBRATION DATA CHANNEL 2							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05045	0.15035	0.19959	STATISTICAL ANALYSIS
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010	+/- ml (Guardband)	0.00129	0.00429	0.00500	
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959	INACCURACY %	0.90008	0.23142	0.20320	
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909	IMPRECISION %	0.19881	0.06671	0.25126	
							Proc. Uncert.(k=2)	0.00021	0.00021	0.00100	
CALIBRATION RESULTS											
ACCURACY							PASS	PASS	PASS		
PRECISION							PASS	PASS	PASS		
Within k=2 95% Confidence							YES	YES	YES		
AS FOUND CALIBRATION DATA CHANNEL 3							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05045	0.15035	0.19959	STATISTICAL ANALYSIS
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010	+/- ml (Guardband)	0.00129	0.00429	0.00500	
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959	INACCURACY %	0.90008	0.23142	0.20320	
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909	IMPRECISION %	0.19881	0.06671	0.25126	
							Proc. Uncert.(k=2)	0.00021	0.00021	0.00100	
CALIBRATION RESULTS											
ACCURACY							PASS	PASS	PASS		
PRECISION							PASS	PASS	PASS		
Within k=2 95% Confidence							YES	YES	YES		
AS FOUND CALIBRATION DATA CHANNEL 4							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05045	0.15035	0.19959	STATISTICAL ANALYSIS
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010	+/- ml (Guardband)	0.00129	0.00429	0.00500	
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959	INACCURACY %	0.90008	0.23142	0.20320	
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909	IMPRECISION %	0.19881	0.06671	0.25126	
							Proc. Uncert.(k=2)	0.00021	0.00021	0.00100	
CALIBRATION RESULTS											
ACCURACY							PASS	PASS	PASS		
PRECISION							PASS	PASS	PASS		
Within k=2 95% Confidence							YES	YES	YES		

Technician: _____

Approved By: _____

Date: _____

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.

ISO 17025:2005



Calibration Certificate #2161.01

YourCompany_City_State_CDS_2.0.8_sample

CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A12345D
NOMINAL SIZE IN (ul)	200
MANUFACTURER	Brand - 8 Ch.
DEPARTMENT	Your Dept.
LOCATION	Your Company
CONTACT	Your Contact
RESTRICTIONS	N/A
CALIBRATION DATE	30-Sep-2008
CALIBRATED BY	DAB
CALIBRATION DUE	30-Mar-09
DATA TYPE	AS FOUND
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70
RELATIVE HUMIDITY IN %	50
BAROMETRIC PRESSURE mmHg	30
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HYGROMETER SERIAL #	1234
CALIBRATION DUE	Jan-2009
BALANCE ID	Brand
BALANCE MODEL	ABC123
BALANCE SERIAL #	ABCD1234
CALIBRATION DUE	Jan-2009
WEIGHT SET SERIAL #	ABC123
CALIBRATION DUE	Aug-2009
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	NO
PREVENTIVE MAINT.	NO

INSTRUMENT STATUS	PASS
--------------------------	-------------

Technician: _____

Approved By: _____

Date: _____

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME ul	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE:AS FOUND
50	3.00	0.00150	3.00		
150	3.00	0.00450	3.00		
200	3.00	0.00600	3.00		

AS FOUND CALIBRATION DATA CHANNEL 4						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909

AS FOUND CALIBRATION DATA CHANNEL 4						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909

AS FOUND CALIBRATION DATA CHANNEL 7						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909

AS FOUND CALIBRATION DATA CHANNEL 4						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.05040	0.05055	0.15000	0.15045	0.19950	0.20010
2	0.05030	0.05045	0.14990	0.15035	0.19900	0.19959
3	0.05020	0.05035	0.14980	0.15025	0.19850	0.19909

STATISTICAL ANALYSIS				
MEAN (ml)	0.05045	0.15035	0.19959	
+/- ml (Guardband)	0.00129	0.00429	0.00500	
INACCURACY %	0.90008	0.23142	0.20320	
IMPRECISION %	0.19881	0.06671	0.25126	
Proc. Uncert.(k=2)	0.00021	0.00021	0.00100	
CALIBRATION RESULTS				
ACCURACY	PASS	PASS	PASS	
PRECISION	PASS	PASS	PASS	
Within k=2 95% Confidence	YES	YES	YES	
STATISTICAL ANALYSIS				
MEAN (ml)	0.05045	0.15035	0.19959	
+/- ml (Guardband)	0.00129	0.00429	0.00500	
INACCURACY %	0.90008	0.23142	0.20320	
IMPRECISION %	0.19881	0.06671	0.25126	
Proc. Uncert.(k=2)	0.00021	0.00021	0.00100	
CALIBRATION RESULTS				
ACCURACY	PASS	PASS	PASS	
PRECISION	PASS	PASS	PASS	
Within k=2 95% Confidence	YES	YES	YES	
STATISTICAL ANALYSIS				
MEAN (ml)	0.05045	0.15035	0.19959	
+/- ml (Guardband)	0.00129	0.00429	0.00500	
INACCURACY %	0.90008	0.23142	0.20320	
IMPRECISION %	0.19881	0.06671	0.25126	
Proc. Uncert.(k=2)	0.00021	0.00021	0.00100	
CALIBRATION RESULTS				
ACCURACY	PASS	PASS	PASS	
PRECISION	PASS	PASS	PASS	
Within k=2 95% Confidence	YES	YES	YES	

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.

ISO 17025:2005



Calibration Certificate #2161.01

YourCompany_City_State_CDS_2.0.8_sample