



610 Jones Ferry Rd. Ste 210, Carrboro, NC 27510
 Ph: 800-253-7064 / Fx: 800-238-2268

Client Information:

Facility Name:	ABC
Street Address:	111 South St.
City, State & Zip:	Raleigh, NC 55555
Contact Name:	John Smith
Telephone:	555-555-5555

Report Information:

Report Number:	
Calibration Date:	9-Aug-2010
Calibration Due:	30-Aug-2011
Payment Reference:	PO #

Specification Source: Calibrate MFG.* ISO* Client*
 * client must supply supporting documentation (SOP's, tolerance charts, etc.)

Location of Service:

<input checked="" type="checkbox"/> On-Site	<i>Cal Lab Locations</i>	
<input type="checkbox"/> Cal Lab	<input type="checkbox"/> Carrboro, NC	610 Jones Ferry Rd., Suite 210 Carrboro, NC 27510
	<input type="checkbox"/> Concord, CA	2520 Stanwell Dr., Suite 100 Concord, CA 94520
	<input type="checkbox"/> Ann Arbor, MI	3600 Green Ct., Suite 400 Ann Arbor, MI 48105

Measurement & Test Equipment Used:

Description	Identification #	Calibration Due
Analytical Balance:	22222	30-Sep-2011
Standard Weight Set:	33333	30-Jul-2011

Service Technician: Bob White

Pipette Service Detail:

#	Identification #	Nominal Size (µl)	Mfg.	# of Channels	Pass Y / N	Parts Replaced / Repairs / Comments
1	A12345D	100	Gilson	1	Y	Replaced Shaft
2	67890	200	VWR	8	Y	Polished Piston Assy.
3	357900054	1000	Matrix	1	Y	
4	Y211567	20	Biohit	1	Y	
5						
6						
7						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						

Calibrate, Inc. hereby certifies that the instrument(s) listed on this Test Report has been tested with standards traceable to the SI through the National Institute of Standards and Technology (N.I.S.T.). The observations recorded are applicable at the time of the observations and apply only to the instrument(s) listed herein. This Test Report shall not be reproduced except in full, in any form, without the expressed written consent of Calibrate, Inc..

Procedure Used: SOP 125-2

Calibrate, Inc. Representative(s):

Date:

Reviewed and Accepted By:

Date:

Bob White

8-Aug-2010