

Calibrate, Inc.

2520 Stanwell Dr. Suite 100
 Concord, CA 94520
 ph.800.253.7064 fx.800.238.2268

CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567C
SIZE (µl)	1200
MANUFACTURER	Matrix 12 CH
LOCATION	CompanyName
DEPARTMENT	QC
CONTACT	Smith, J.
RESTRICTION	GMP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS FOUND
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	NO
PREVENTIVE MAINT.	NO
INSTRUMENT STATUS	FAIL

Technician: _____

Approved By: _____

Date: 01-Nov-2010

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME ul	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE:AS FOUND
300	3.00	0.00900	3.00		
1200	3.00	0.03600	3.00		

CDS 2.1.28

copyright 1997-2011

AS FOUND CALIBRATION DATA CHANNEL 1						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.27639	1.08784	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	+/- ml (Guardband)	0.00133	0.03146		
2	0.27830	0.27913	1.08430	1.08753	INACCURACY %	7.87045	9.34706		
3	0.27120	0.27201	1.08250	1.08573	IMPRECISION %	1.38675	0.20883		
						Proc. Uncert.(k=2)	0.00767		0.00454
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
AS FOUND CALIBRATION DATA CHANNEL 2						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00133	0.03146	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	INACCURACY %	7.87045	9.34706		
2	0.27830	0.27913	1.08430	1.08753	IMPRECISION %	1.38675	0.20883		
3	0.27120	0.27201	1.08250	1.08573	Proc. Uncert.(k=2)	0.00767	0.00454		
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
AS FOUND CALIBRATION DATA CHANNEL 3						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00133	0.03146	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	INACCURACY %	7.87045	9.34706		
2	0.27830	0.27913	1.08430	1.08753	IMPRECISION %	1.38675	0.20883		
3	0.27120	0.27201	1.08250	1.08573	Proc. Uncert.(k=2)	0.00767	0.00454		
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
AS FOUND CALIBRATION DATA CHANNEL 4						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00133	0.03146	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	INACCURACY %	7.87045	9.34706		
2	0.27830	0.27913	1.08430	1.08753	IMPRECISION %	1.38675	0.20883		
3	0.27120	0.27201	1.08250	1.08573	Proc. Uncert.(k=2)	0.00767	0.00454		
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



Calibrate, Inc.

2520 Stanwell Dr. Suite 100
 Concord, CA 94520
 ph.800.253.7064 fx.800.238.2268

CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567C
SIZE (µl)	1200
MANUFACTURER	Matrix 12 CH
LOCATION	CompanyName
DEPARTMENT	QC
CONTACT	Smith, J.
RESTRICTION	GMP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS FOUND
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	NO
PREVENTIVE MAINT.	NO
INSTRUMENT STATUS	
	FAIL

Technician: _____

Approved By: _____

Date: 01-Nov-2010

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME µl	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE:AS FOUND
300	3.00	0.00900	3.00		
1200	3.00	0.03600	3.00		

CDS 2.1.28

copyright 1997-2011

AS FOUND CALIBRATION DATA CHANNEL 5						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.27639	1.08784	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	+/- ml (Guardband)	0.00133	0.03146		
2	0.27830	0.27913	1.08430	1.08753	INACCURACY %	7.87045	9.34706		
3	0.27120	0.27201	1.08250	1.08573	IMPRECISION %	1.38675	0.20883		
						Proc. Uncert.(k=2)	0.00767		0.00454
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
AS FOUND CALIBRATION DATA CHANNEL 6						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00133	0.03146	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	INACCURACY %	7.87045	9.34706		
2	0.27830	0.27913	1.08430	1.08753	IMPRECISION %	1.38675	0.20883		
3	0.27120	0.27201	1.08250	1.08573	Proc. Uncert.(k=2)	0.00767	0.00454		
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
						+/- ml (Guardband)	0.00133		0.03146
AS FOUND CALIBRATION DATA CHANNEL 7						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	INACCURACY %	7.87045	9.34706	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	IMPRECISION %	1.38675	0.20883		
2	0.27830	0.27913	1.08430	1.08753	Proc. Uncert.(k=2)	0.00767	0.00454		
3	0.27120	0.27201	1.08250	1.08573					
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
						+/- ml (Guardband)	0.00133		0.03146
AS FOUND CALIBRATION DATA CHANNEL 8						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	INACCURACY %	7.87045	9.34706	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	IMPRECISION %	1.38675	0.20883		
2	0.27830	0.27913	1.08430	1.08753	Proc. Uncert.(k=2)	0.00767	0.00454		
3	0.27120	0.27201	1.08250	1.08573					
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
						+/- ml (Guardband)	0.00133		0.03146

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



Calibrate, Inc.

2520 Stanwell Dr. Suite 100
 Concord, CA 94520
 ph.800.253.7064 fx.800.238.2268

CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567C
SIZE (µl)	1200
MANUFACTURER	Matrix 12 CH
LOCATION	CompanyName
DEPARTMENT	QC
CONTACT	Smith, J.
RESTRICTION	GMP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS FOUND
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	NO
PREVENTIVE MAINT.	NO
INSTRUMENT STATUS	
	FAIL

Technician: _____

Approved By: _____

Date: 01-Nov-2010

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME ul	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE:AS FOUND
300	3.00	0.00900	3.00		
1200	3.00	0.03600	3.00		

CDS 2.1.28

copyright 1997-2011

AS FOUND CALIBRATION DATA CHANNEL 9						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.27639	1.08784	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	+/- ml (Guardband)	0.00133	0.03146		
2	0.27830	0.27913	1.08430	1.08753	INACCURACY %	7.87045	9.34706		
3	0.27120	0.27201	1.08250	1.08573	IMPRECISION %	1.38675	0.20883		
						Proc. Uncert.(k=2)	0.00767		0.00454
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
AS FOUND CALIBRATION DATA CHANNEL 10						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00133	0.03146	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	INACCURACY %	7.87045	9.34706		
2	0.27830	0.27913	1.08430	1.08753	IMPRECISION %	1.38675	0.20883		
3	0.27120	0.27201	1.08250	1.08573	Proc. Uncert.(k=2)	0.00767	0.00454		
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
AS FOUND CALIBRATION DATA CHANNEL 11						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00133	0.03146	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	INACCURACY %	7.87045	9.34706		
2	0.27830	0.27913	1.08430	1.08753	IMPRECISION %	1.38675	0.20883		
3	0.27120	0.27201	1.08250	1.08573	Proc. Uncert.(k=2)	0.00767	0.00454		
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784
AS FOUND CALIBRATION DATA CHANNEL 12						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00133	0.03146	STATISTICAL ANALYSIS	
1	0.27720	0.27803	1.08700	1.09024	INACCURACY %	7.87045	9.34706		
2	0.27830	0.27913	1.08430	1.08753	IMPRECISION %	1.38675	0.20883		
3	0.27120	0.27201	1.08250	1.08573	Proc. Uncert.(k=2)	0.00767	0.00454		
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	NO		NO
						MEAN (ml)	0.27639		1.08784

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567C
SIZE (µl)	1200
MANUFACTURER	Matrix 12 CH
LOCATION	CompanyName
DEPARTMENT	QC
CONTACT	Smith, J.
RESTRICTION	GMP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS LEFT
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	YES
PREVENTIVE MAINT.	YES
INSTRUMENT STATUS	
PASS	

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME µl	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE: AS LEFT
300	3.00	0.00900	3.00		
900	3.00	0.02700	3.00		
1200	3.00	0.03600	3.00		

AS LEFT CALIBRATION DATA CHANNEL 1						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 2						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 3						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 4						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

CALIBRATION RESULTS					STATISTICAL ANALYSIS
MEAN (ml)	0.30193	0.90285	1.20465		
+/- ml (Guardband)	0.00799	0.02569	0.03329		
INACCURACY %	0.64376	0.31686	0.38744		
IMPRECISION %	0.16720	0.07228	0.11243		
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271		
CALIBRATION RESULTS					STATISTICAL ANALYSIS
ACCURACY	PASS	PASS	PASS		
PRECISION	PASS	PASS	PASS		
Within k=2 95% Confidence	YES	YES	YES		
MEAN (ml)	0.30193	0.90285	1.20465		
+/- ml (Guardband)	0.00799	0.02569	0.03329		
INACCURACY %	0.64376	0.31686	0.38744		
IMPRECISION %	0.16720	0.07228	0.11243		
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271		
CALIBRATION RESULTS					STATISTICAL ANALYSIS
ACCURACY	PASS	PASS	PASS		
PRECISION	PASS	PASS	PASS		
Within k=2 95% Confidence	YES	YES	YES		
MEAN (ml)	0.30193	0.90285	1.20465		
+/- ml (Guardband)	0.00799	0.02569	0.03329		
INACCURACY %	0.64376	0.31686	0.38744		
IMPRECISION %	0.16720	0.07228	0.11243		
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271		
CALIBRATION RESULTS					STATISTICAL ANALYSIS
ACCURACY	PASS	PASS	PASS		
PRECISION	PASS	PASS	PASS		
Within k=2 95% Confidence	YES	YES	YES		

Technician: _____

Approved By: _____

Date: 01-Nov-2010

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567C
SIZE (µl)	1200
MANUFACTURER	Matrix 12 CH
LOCATION	CompanyName
DEPARTMENT	QC
CONTACT	Smith, J.
RESTRICTION	GMP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS LEFT
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	YES
PREVENTIVE MAINT.	YES
INSTRUMENT STATUS	
PASS	

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME µl	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE: AS LEFT
300	3.00	0.00900	3.00		
900	3.00	0.02700	3.00		
1200	3.00	0.03600	3.00		

AS LEFT CALIBRATION DATA CHANNEL 5						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 6						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 7						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 8						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

CALIBRATION RESULTS					STATISTICAL ANALYSIS
MEAN (ml)	0.30193	0.90285	1.20465		
+/- ml (Guardband)	0.00799	0.02569	0.03329		
INACCURACY %	0.64376	0.31686	0.38744		
IMPRECISION %	0.16720	0.07228	0.11243		
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271		
CALIBRATION RESULTS					STATISTICAL ANALYSIS
ACCURACY	PASS	PASS	PASS		
PRECISION	PASS	PASS	PASS		
Within k=2 95% Confidence	YES	YES	YES		
MEAN (ml)	0.30193	0.90285	1.20465		
+/- ml (Guardband)	0.00799	0.02569	0.03329		
INACCURACY %	0.64376	0.31686	0.38744		
IMPRECISION %	0.16720	0.07228	0.11243		
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271		
CALIBRATION RESULTS					STATISTICAL ANALYSIS
ACCURACY	PASS	PASS	PASS		
PRECISION	PASS	PASS	PASS		
Within k=2 95% Confidence	YES	YES	YES		
MEAN (ml)	0.30193	0.90285	1.20465		
+/- ml (Guardband)	0.00799	0.02569	0.03329		
INACCURACY %	0.64376	0.31686	0.38744		
IMPRECISION %	0.16720	0.07228	0.11243		
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271		
CALIBRATION RESULTS					STATISTICAL ANALYSIS
ACCURACY	PASS	PASS	PASS		
PRECISION	PASS	PASS	PASS		
Within k=2 95% Confidence	YES	YES	YES		

CDS 2.1.28
 copyright 1997-2011

Technician: _____

Approved By: _____

Date: 01-Nov-2010

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567C
SIZE (µl)	1200
MANUFACTURER	Matrix 12 CH
LOCATION	CompanyName
DEPARTMENT	QC
CONTACT	Smith, J.
RESTRICTION	GMP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS LEFT
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	YES
PREVENTIVE MAINT.	YES
INSTRUMENT STATUS	
PASS	

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME µl	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE: AS LEFT
300	3.00	0.00900	3.00		
900	3.00	0.02700	3.00		
1200	3.00	0.03600	3.00		

AS LEFT CALIBRATION DATA CHANNEL 9						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 10						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 11						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

AS LEFT CALIBRATION DATA CHANNEL 12						
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)
1	0.30110	0.30200	0.90020	0.90289	1.20240	1.20599
2	0.30050	0.30140	0.90080	0.90349	1.20110	1.20468
3	0.30150	0.30240	0.89950	0.90218	1.19970	1.20328

CALIBRATION RESULTS				
MEAN (ml)	0.30193	0.90285	1.20465	STATISTICAL ANALYSIS
+/- ml (Guardband)	0.00799	0.02569	0.03329	
INACCURACY %	0.64376	0.31686	0.38744	
IMPRECISION %	0.16720	0.07228	0.11243	
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271	
ACCURACY	PASS	PASS	PASS	
PRECISION	PASS	PASS	PASS	
Within k=2 95% Confidence	YES	YES	YES	
MEAN (ml)	0.30193	0.90285	1.20465	
+/- ml (Guardband)	0.00799	0.02569	0.03329	
CALIBRATION RESULTS				
MEAN (ml)	0.30193	0.90285	1.20465	STATISTICAL ANALYSIS
+/- ml (Guardband)	0.00799	0.02569	0.03329	
INACCURACY %	0.64376	0.31686	0.38744	
IMPRECISION %	0.16720	0.07228	0.11243	
Proc. Uncert.(k=2)	0.00101	0.00131	0.00271	
ACCURACY	PASS	PASS	PASS	
PRECISION	PASS	PASS	PASS	
Within k=2 95% Confidence	YES	YES	YES	
MEAN (ml)	0.30193	0.90285	1.20465	
+/- ml (Guardband)	0.00799	0.02569	0.03329	

Technician: _____

Approved By: _____

Date: 01-Nov-2010

Note:

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.

