

Calibrate, Inc.

2520 Stanwell Dr. Suite 100
 Concord, CA 94520
 ph.800.253.7064 fx.800.238.2268

CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567B
SIZE (µl)	200
MANUFACTURER	Biohit 8 CH
LOCATION	CompanyName
DEPARTMENT	QA
CONTACT	Smith, J.
RESTRICTION	GLP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS FOUND
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	NO
PREVENTIVE MAINT.	NO

INSTRUMENT STATUS	FAIL
--------------------------	-------------

Technician: _____

Approved By: _____

Date: 01-Nov-2010

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME ul	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE:AS FOUND
50	3.00	0.00150	3.00		
200	3.00	0.00600	3.00		

CDS 2.1.28

copyright 1997-2011

AS FOUND CALIBRATION DATA CHANNEL 1						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.04580	0.16656		
1	0.04620	0.04634	0.16800	0.16850	+/- ml (Guardband)	0.00039	0.00013		
2	0.04510	0.04523	0.16750	0.16800	INACCURACY %	8.39423	16.71899		
3	0.04570	0.04584	0.16270	0.16319	IMPRECISION %	1.20604	1.76213		
						Proc. Uncert.(k=2)	0.00111		0.00587
						CALIBRATION RESULTS			
						ACCURACY	FAIL		FAIL
						PRECISION	PASS	PASS	
						Within k=2 95% Confidence			
						MEAN (ml)	0.05018	0.20063	
						+/- ml (Guardband)	0.00088	0.00489	
						INACCURACY %	0.36515	0.31500	
						IMPRECISION %	0.61060	0.27533	
						Proc. Uncert.(k=2)	0.00062	0.00111	
						CALIBRATION RESULTS			
						ACCURACY	PASS	PASS	
						PRECISION	PASS	PASS	
						Within k=2 95% Confidence			
						MEAN (ml)	0.05018	0.20063	
						+/- ml (Guardband)	0.00088	0.00489	
						INACCURACY %	0.36515	0.31500	
						IMPRECISION %	0.61060	0.27533	
						Proc. Uncert.(k=2)	0.00062	0.00111	
						CALIBRATION RESULTS			
						ACCURACY	PASS	PASS	
						PRECISION	PASS	PASS	
						Within k=2 95% Confidence			
						MEAN (ml)	0.04223	0.12654	
						+/- ml (Guardband)	0.00070	0.00210	
						INACCURACY %	15.54884	36.72850	
						IMPRECISION %	0.95012	1.53964	
						Proc. Uncert.(k=2)	0.00080	0.00390	
						CALIBRATION RESULTS			
						ACCURACY	FAIL	FAIL	
						PRECISION	PASS	PASS	
						Within k=2 95% Confidence			

Note: Channel #1 and #4 are leaking.

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



Calibrate, Inc.

2520 Stanwell Dr. Suite 100
 Concord, CA 94520
 ph.800.253.7064 fx.800.238.2268

CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567B
SIZE (µl)	200
MANUFACTURER	Biohit 8 CH
LOCATION	CompanyName
DEPARTMENT	QA
CONTACT	Smith, J.
RESTRICTION	GLP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS FOUND
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	NO
ADJUST CALIBRATION	NO
PREVENTIVE MAINT.	NO

INSTRUMENT STATUS	FAIL
--------------------------	-------------

Technician: _____

Approved By: _____

Date: 01-Nov-2010

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME ul	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE:AS FOUND
50	3.00	0.00150	3.00		
200	3.00	0.00600	3.00		

CDS 2.1.28

AS FOUND CALIBRATION DATA CHANNEL 5						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.20063	STATISTICAL ANALYSIS	
1	0.05010	0.05025	0.20030	0.20090	+/- ml (Guardband)	0.00088	0.00489		
2	0.04970	0.04985	0.19940	0.19999	INACCURACY %	0.36515	0.31500		
3	0.05030	0.05045	0.20040	0.20100	IMPRECISION %	0.61060	0.27533		
						Proc. Uncert.(k=2)	0.00062		0.00111
						CALIBRATION RESULTS			
						ACCURACY	PASS		PASS
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	YES		YES
						MEAN (ml)	0.05018		0.20063
AS FOUND CALIBRATION DATA CHANNEL 6						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	+/- ml (Guardband)	0.00088	0.00489	STATISTICAL ANALYSIS	
1	0.05010	0.05025	0.20030	0.20090	INACCURACY %	0.36515	0.31500		
2	0.04970	0.04985	0.19940	0.19999	IMPRECISION %	0.61060	0.27533		
3	0.05030	0.05045	0.20040	0.20100	Proc. Uncert.(k=2)	0.00062	0.00111		
						CALIBRATION RESULTS			
						ACCURACY	PASS		PASS
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	YES		YES
						MEAN (ml)	0.05018		0.20063
						+/- ml (Guardband)	0.00088		0.00489
AS FOUND CALIBRATION DATA CHANNEL 7						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	INACCURACY %	0.36515	0.31500	STATISTICAL ANALYSIS	
1	0.05010	0.05025	0.20030	0.20090	IMPRECISION %	0.61060	0.27533		
2	0.04970	0.04985	0.19940	0.19999	Proc. Uncert.(k=2)	0.00062	0.00111		
3	0.05030	0.05045	0.20040	0.20100					
						CALIBRATION RESULTS			
						ACCURACY	PASS		PASS
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	YES		YES
						MEAN (ml)	0.05018		0.20063
						+/- ml (Guardband)	0.00088		0.00489
AS FOUND CALIBRATION DATA CHANNEL 8						STATISTICAL ANALYSIS			
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	INACCURACY %	0.36515	0.31500	STATISTICAL ANALYSIS	
1	0.05010	0.05025	0.20030	0.20090	IMPRECISION %	0.61060	0.27533		
2	0.04970	0.04985	0.19940	0.19999	Proc. Uncert.(k=2)	0.00062	0.00111		
3	0.05030	0.05045	0.20040	0.20100					
						CALIBRATION RESULTS			
						ACCURACY	PASS		PASS
						PRECISION	PASS		PASS
						Within k=2 95% Confidence	YES		YES
						MEAN (ml)	0.05018		0.20063
						+/- ml (Guardband)	0.00088		0.00489

Note: Channel #1 and #4 are leaking.

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567B
SIZE (µl)	200
MANUFACTURER	Biohit 8 CH
LOCATION	CompanyName
DEPARTMENT	QA
CONTACT	Smith, J.
RESTRICTION	GLP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS LEFT
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	YES
ADJUST CALIBRATION	YES
PREVENTIVE MAINT.	YES
INSTRUMENT STATUS	
PASS	

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME µl	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE: AS LEFT
50	3.00	0.00150	3.00		
150	3.00	0.00450	3.00		
200	3.00	0.00600	3.00		

AS LEFT CALIBRATION DATA CHANNEL 1							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	STATISTICAL ANALYSIS
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	
AS LEFT CALIBRATION DATA CHANNEL 2							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	
AS LEFT CALIBRATION DATA CHANNEL 3							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	STATISTICAL ANALYSIS
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	
AS LEFT CALIBRATION DATA CHANNEL 4							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	

CDS 2.1.28

copyright 1997-2011

Technician: _____

Approved By: _____

Date: 01-Nov-2010

Note: Channel #1 and #4 repaired.

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.



CALIBRATION CERTIFICATE

PIPETTE INFORMATION	
SERIAL #	A1234567B
SIZE (µl)	200
MANUFACTURER	Biohit 8 CH
LOCATION	CompanyName
DEPARTMENT	QA
CONTACT	Smith, J.
RESTRICTION	GLP
CALIBRATION DATE	01-Nov-2010
CALIBRATED BY	PS
CALIBRATION DUE	01-May-2011
DATA TYPE	AS LEFT
ENVIRONMENTAL CONDITIONS	
TEMPERATURE IN F	70.00
RELATIVE HUMIDITY IN %	50.00
BAROMETRIC PRESSURE INCHES	30.00
AIR DENSITY	0.0012030425621
Z FACTOR	1.0029828772
MEASUREMENT & TEST EQUIP. INFO	
HERMO-HYGROMETER SERIAL #	90917310
CALIBRATION DUE	30-Nov-2010
BALANCE ID	METTLER
BALANCE MODEL	AG245
BALANCE SERIAL #	1117023184
CALIBRATION DUE	28-Feb-2011
WEIGHT SET SERIAL #	1MN2
CALIBRATION DUE	31-Aug-2012
PARTS REPLACED	
SEAL	NO
O RING	NO
PISTON	NO
SHAFT	NO
TENSION RING	NO
REPAIR	YES
ADJUST CALIBRATION	YES
PREVENTIVE MAINT.	YES
INSTRUMENT STATUS	
PASS	

TOLERANCE LIMITS				SERVICE INFORMATION	
VOLUME µl	INACCURACY %	+/- ml	IMPRECISION %	SOP Followed	Level 6 SOP 125-6 DATA TYPE: AS LEFT
50	3.00	0.00150	3.00		
150	3.00	0.00450	3.00		
200	3.00	0.00600	3.00		

CDS 2.1.28

copyright 1997-2011

AS LEFT CALIBRATION DATA CHANNEL 5							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	STATISTICAL ANALYSIS
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	STATISTICAL ANALYSIS
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	
AS LEFT CALIBRATION DATA CHANNEL 6							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	STATISTICAL ANALYSIS
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	STATISTICAL ANALYSIS
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	
AS LEFT CALIBRATION DATA CHANNEL 7							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	STATISTICAL ANALYSIS
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	STATISTICAL ANALYSIS
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	
AS LEFT CALIBRATION DATA CHANNEL 8							STATISTICAL ANALYSIS				
SAMPLE	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	WEIGHT(g)	VOLUME(ml)	MEAN (ml)	0.05018	0.15058	0.20133	STATISTICAL ANALYSIS
1	0.05020	0.05035	0.15020	0.15065	0.20050	0.20110	+/- ml (Guardband)	0.00108	0.00369	0.00535	
2	0.04980	0.04995	0.14970	0.15015	0.20110	0.20170	INACCURACY %	0.36515	0.38744	0.66605	
3	0.05010	0.05025	0.15050	0.15095	0.20060	0.20120	IMPRECISION %	0.41606	0.26919	0.16014	
							Proc. Uncert.(k=2)	0.00042	0.00081	0.00065	STATISTICAL ANALYSIS
CALIBRATION RESULTS							ACCURACY	PASS	PASS	PASS	
							PRECISION	PASS	PASS	PASS	
							Within k=2 95% Confidence	YES	YES	YES	

Note: Channel #1 and #4 repaired.

Technician: _____

Approved By: _____

Date: 01-Nov-2010

This calibration was conducted using standards traceable to SI through N.I.S.T. This document may not be reproduced except in full.

